


Service Repair Sheet SRS69353

Contact Name Karen Carvell
 Company/ Hospital Name London Bridge Hospital
 Department
 Position UK HCA Site Administrator
 Direct Phone 07977700502
 General Phone 02035539022
 Opera Account 00003004
 Email karen.carvell@ergeagroup.com
 Order Number
 Date Received 11/Feb/2026
 Booked in By Robert Connor
 Main Company Viamed
 Type Return Quote

VIAMEDclean	
Goods In Only	<input checked="" type="checkbox"/>
Decontamination certificate provided by customer	
Cleaned by Viamed, if no declaration certificate from customer	
Signed: _____	
Date: _____	
Goods Out Only	
Cleaned by Viamed before returning to customer	
Signed: _____	
Date: _____	

Notes 30/Jan/2026 Kate Griffiths
 30/Jan/2026 Kate Griffiths
 will send in 2 x MX300. s/n: 267368 and 267366. They have checked with another sensor and another cable and still get sensor error message. Have asked for cert of decont/batteries removed/to send in with cable and sensor/SRS on all paperwork
 04/Feb/2026 Kate Griffiths
 PO S01-PO064541-1 recd before device
 11/Feb/2026 Robert Connor
 Received 2 x MX s/n 267366, 267368, 2 x cable s/n 0001947, 0003874, 1 x R17V s/n V100657

Ready For quote CARVELL 16.2.26.
 Repair Complete Signed CARVELL 19.2.26.

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN38536	Oxygen Monitor	0111237	267366	22222
SRN38537	Oxygen Monitor	0111237	267368	
SRN38538	Misc. Cable	0131202	0001947	
SRN38539	Misc. Cable	0131202	0003874	
SRN38540	Oxygen Sensor	0110018	V100657	

0180005 x 2 @ £50 each
 S/N's, SRS, SRN's

0140025 x 2 FOC
 S/N's, SRS, SRN
 NO FAULT FOUND.

0140012 x 1 FOC
 old S/N, SRS, SRN

0110018 x 1 £43.10.
 NEW NEW S/N, SRS, SRN
 UPS x 1 @ £12