

Service Repair Sheet SRS69334

Contact Name	Sean Bunn
Company/ Hospital Name	Luton & Dunstable Hospital
Department	Clinical Engineering
Position	
Direct Phone	01582718369
General Phone	01582497042
Opera Account	00003490
Email	Sean.Bunn@ldh.nhs.uk
Order Number	
Date Received	15/Jan/2026
Booked in By	Robert Connor
Main Company	Viamed
Type Return	Quote

VIAMEDclean	
Goods In Only	
Decontamination certificate provided by customer	<input checked="" type="checkbox"/>
Cleaned by Viamed, if no declaration certificate from customer	<input checked="" type="checkbox"/>
Signed: _____	
Date: _____	
Goods Out Only	
Cleaned by Viamed before returning to customer	
Signed: _____	
Date: _____	

Notes 14/Jan/2026 Catrin Hird

14/Jan/2026 Catrin Hird

Customer returning VM-2500-S, S/N: A1711160013

The unit needs recalibration, but they usually do this themselves but when checking the percentage of the CO2 was off, even after zeroing. Please review unit and confirm with customer before proceeding with any calibration or work.

15/Jan/2026 Robert Connor

Received 1 x VM-2500-S s/n A1711160013, with silicone cover and 4 x AA battery.

Ready For quote

CGreen 20.2.26

Repair Complete Signed _____

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN38504	Capnograph	4410520	A1711160013	2

4440000 x 45mins @ £95 per hr.
S/N, SRS, SRN

- Includes our time + Blueprints.

4480520 x 1 @ £160
SRS, SRN

4430201 x 1 @ £15.40
SRS, SRN

4430050 x 1 @ £2.85.
SRS, SRN

UPS x 1 @ £12.