

Service Repair Sheet 69314
Contact Name Gurkaran Johal
Company/ Hospital Name Royal London Hospital
Direct Phone 02035942240
General Phone 02073777000
Opera Account 00003030
Email gurkaran.johal@nhs.net
Order Number 41245344
Date Received 11/Feb/2026
Booked in By Robert Connor

Repair	Ref	S/N	Equipment Type	Complete - Repaired
SRN38534	1410000	PR065A17	Foetal Simulator	Time :0 Hour(s)
Parts Replaced				
	Qty: 1	1480000		
	Qty: 1	1430309		

V1000 has had transducer interface cushion replaced and calibration has been checked.
