

Service Repair Sheet 69368
Contact Name Ted Gibbs
Company/ Hospital Name Southern Calibration Laboratories
Direct Phone 01489790296
General Phone 01489790296
Opera Account CID22311
Email tedg@southcal.co.uk
Order Number SCL26/0175/5015
Date Received 23/Feb/2026
Booked in By Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN38550	1410000	PR0545A14	Foetal Simulator	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
	Qty: 1	1480000		
	Qty: 1	1430309		

V1000 has had transducer interface cushion replaced and calibration has been checked.
