

Service Repair Sheet 69321  
Contact Name Damien Mahendrarajah  
Company/ Hospital Name Royal Brompton and Harefield  
Department Clinical Engineering  
Position Clinical Engineering Technologist  
Direct Phone 02073528121  
General Phone 020 7352 8121  
Opera Account 00002790  
Email damien.mahendrarajah@nhs.net  
Order Number RJ1-1464936  
Date Received 08/Jan/2026  
Booked in By Robert Connor

---

Repair	Ref	S/N	Equipment Type	
SRN38511	0111276	100107	Oxygen Monitor	Complete - Repaired Time :1.5 Hour(s)
Parts Replaced		Qty: 1.5	0140000	

The MySign O Oxygen monitor has been repaired. The unit now charges and operates correctly.

---