

Certificate of Calibration

Certificate number:

Tom Thumb Infant Resuscitator.

Part Number:

Serial number:

This certificate reports recorded values for the instrument specified above.

The output rates were measured in accordance with Tom Thumb service manual.

Equipment used: Digitron Digital Manometer P200 (CE)
Digitron Digital Manometer P200 (CE)
Flow Meter (CE130)
Snoop liquid

Test Equipment	Test	Specification	Reading	Pass/Fail
Snoop liquid	Check all ports / connections for leaks	No bubbling	N/A	
Digital Manometer (CE***)	Adjustable Valve: @ 10 Lpm	Minimum: <8 cmH ₂ O		
Digital Manometer (CE***)	Adjustable Valve: @ 10 Lpm	Maximum: 43±1 cmH ₂ O		
Digital Manometer (CE***)	Precision Valve: @ 10 Lpm	Maximum: 45±1 cmH ₂ O		
Digital Manometer (CE***)	Pressure Gauge Test @30cmH ₂ O	Pressure: 30cmH ₂ O ±0.5		

Name:

Date:

Signed: _____

