

Service Repair Sheet 69352  
Contact Name Andrew Ruddick  
Company/ Hospital Name Sunderland Royal Hospital  
Department Clinical Engineering Department  
Position Clinical Engineer  
Direct Phone 01915699933  
General Phone 01915656256  
Opera Account 00004992  
Email andrew.ruddick@nhs.net  
Order Number RLNP400137048  
Date Received 02/Mar/2026  
Booked in By Robert Connor

---

Repair	Ref	S/N	Equipment Type	
SRN38556	1410000	GE71743924	Foetal Simulator	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
		Qty: 1	1480000	
		Qty: 1	1430309	

V1000 has had transducer interface cushion replaced and calibration has been checked.

---