

Service Repair Sheet 69375

Contact Name Marlene Laverty  
Company/ Hospital Name Queen Elizabeth University Hospital  
Department Medical Physics  
Position Assistant Higher Level  
Direct Phone 01414523282  
General Phone 01412011100  
Opera Account 00001853  
Email marlene.laverty2@nhs.scot  
Order Number GMPS16325001  
Date Received 26/Feb/2026  
Booked in By Robert Connor

---

Repair	Ref	S/N	Equipment Type	
SRN38555	1410000	PR02041A10	Foetal Simulator	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
		Qty: 1	1480000	
		Qty: 1	1430309	

V1000 has had transducer interface cushion replaced and calibration has been checked.

---