

Service Repair Sheet 69378

Contact Name Ross Bowden
Company/ Hospital Name Cardiff and Vale UHB
Department Clinical Engineering
Direct Phone 02921845678
General Phone 029 2090 3794
Opera Account 00000092
Email Ross.Bowden@wales.nhs.uk
Order Number 726865964
Date Received 02/Mar/2026
Booked in By Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN38557	1410000	PR02000A10	Foetal Simulator	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
		Qty: 1	1480000	
		Qty: 1	1430309	

V1000 has had transducer interface cushion replaced and calibration has been checked.
