

Service Repair Sheet 69374

Contact Name Balraj Raghavanunni  
Company/ Hospital Name University College Hospital  
Department Medical Physics and Biomedical Engineering  
Position Clinical Engineering Technologist  
Direct Phone 02034474848  
General Phone 02034474848  
Opera Account 00003420  
Email balraj.raghavanunni@nhs.net  
Order Number RRVN400427395  
Date Received 26/Feb/2026  
Booked in By Robert Connor

---

Repair	Ref	S/N	Equipment Type	
SRN38552	1410000	PR03485A13	Foetal Simulator	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
		Qty: 1	1480000	
		Qty: 1	1430309	

V1000 has had transducer interface cushion replaced and calibration has been checked.

---