

Service Repair Sheet 69384

Contact Name Christopher Courage
Company/ Hospital Name Royal Cornwall Hospital
Department Department of Clinical Technology
Position Entry-Level Technician
Direct Phone 01872252488
Email christopher.courage@nhs.net
Date Received 19/Mar/2026
Booked in By Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN38672	1410000	GE71743920	Foetal Simulator	Complete - Repaired Time :0 Hour(s)

Parts Replaced

Qty: 1 1480000

Qty: 1 1430309

V1000 has had transducer interface cushion replaced and calibration has been checked.
