

Service Repair Sheet 69354
Contact Name Dave Bell
Company/ Hospital Name North Tyneside General Hospital
Department Medical Engineering
Position Medical Engineering Area Team Leader
Direct Phone 03448118111
General Phone 01912934233
Opera Account 00003859
Email david.bell2@NHCT.NHS.UK
Order Number RTFE400475703
Date Received 25/Mar/2026
Booked in By Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN38685	1410000	PR03124A10	Foetal Simulator	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
		Qty: 1	1480000	
		Qty: 1	1430309	

V1000 has had transducer interface cushion replaced and calibration has been checked.
