

Service Repair Sheet SRS69301

Contact Name

Adam Wigfield

Company/ Hospital Name

Inspiration Healthcare Ltd

Department

Position

Direct Phone

General Phone

03301750000

Opera Account

00012240

Email

Order Number

Date Received

19/Mar/2026

Booked in By

Robert Connor

Main Company

Viamed

Type Return

Bring To Attention

VIAMEDclean	
Goods In Only Decontamination certificate provided by customer	<input type="checkbox"/>
Cleaned by Viamed, if no declaration certificate from customer Signed: _____ Date: _____	<input type="checkbox"/>
Goods Out Only Cleaned by Viamed before returning to customer Signed: _____ Date: _____	

Notes They have been informed by their customer that 1 x R-22MEDV 0110021 s/n V121884 failed out of box, the sensor is at 0V no matter the actual Ox concentration.

25/Nov/2025 Kate Griffiths

They have been informed by their customer that 1 x R-22MEDV 0110021 s/n V121884 failed out of box, the sensor is at 0V no matter the actual Ox concentration.

19/Mar/2026 Robert Connor

Received 1 x R22MEDV s/n V121884.

Ready For quote

CGreen 19-3-26

GBP Distrib.

Repair Complete Signed

CGreen 26-3-26

ok to keep as warranty as agreed with SN.
CG 26-3-26

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN38673	Oxygen Sensor	0110021	V121884	N

0140012 x 1 FOC
old S/N, SRS, SRN

0110021 x 1 ~~FOC~~ FOC.
SRS 69301, SRN 38673
↑ new S/N

~~UPS~~ x 1 ~~E12 UPS~~ ~~EXW~~ FOC