

#### **Post Market Surveillance Review**

# 4310001 Flowsensor SpiroTrue A. Box of 5

First sold: 12th January 2009

General sales suspended 18th December 2013 due to MHRA stance on single use designation.

Total sales to date: 5,906 boxes, across 34 countries

Total failures: 6 boxes
Total returns: 6 boxes

Total customer failures: 0.102%

# Liverpool Heart & Chest Hospital SRS63066, received 1 box 10<sup>th</sup> October 2011

Failed calibration. Also reported sensors generally have shorter life than Draeger OEM sensors.

**Result:** Customer lost, approximately 100 boxes per year. Resumed using Draeger OEM sensors, will try and regain the sales with Flowsensor SpiroTrue A.

# Dorset County Hospital SRS63869, received 1 box 3<sup>rd</sup> June 2013

Failed calibration.

Result: Customer lost, approximately 12 boxes per year. Will try and regain the sales with Flowsensor SpiroTrue A.

# Papworth Hospital SRS64002, received 4 boxes 11th September 2013

Failed calibration when used on Evita 4s.

**Result:** Customer lost, approximately 120 boxes per year. They have tried both SpiroTrue A and SpiroTrue A (PC). Will try and regain the sales with Flowsensor SpiroTrue A.

### Applications:

Manufacturer	Description	
Acutronic	1601+	Flow Sensor
Acutronic	SpiroQuant A	Flow Sensor
Asmuth GmbH Medinzintechnik	AS432+	Flow Sensor
Asmuth GmbH Medinzintechnik	AS8403735	Flow Sensor
Asmuth GmbH Medinzintechnik	SpiroTrue A	Flow Sensor
Draeger	Spirolog 84 03 735	Flow Sensor
Draeger	8403735	Flow Sensor
Draeger	Anemone	Monitor
Draeger	Apollo	Anaesthesia System
Draeger	AV1	Anaesthesia System
Draeger	Cato	Anaesthesia System
Draeger	Cicero	Anaesthesia System
Draeger	D-9118-2009	Flow Sensor
Draeger	Dura	Ventilator
Draeger	Eva	Ventilator
Draeger	Evita	Ventilator

Evita 4 Ventilator Draeger Evita XL Draeger Ventilator Fabius GS Draeger Anaesthesia System Infinity ID (NOTE Viamed does not have RFID) Flow Sensor Draeger Draeger Infinity V500 (NOTE Viamed does not have RFID) Anaesthesia System Anaesthesia System Draeger Julian Draeger MT-2086-2007 (NOTE Viamed does not have RFID) Flow Sensor PM 8030 Monitor Draeger PM 8035 Monitor Draeger PM 8050 Monitor Draeger PM 8060 Monitor Draeger Draeger **Primus** Anaesthesia System Draeger Savina Ventilator Flow Sensor Draeger Spirolog **Spirolog 8403735** Flow Sensor Draeger Sulla Draeger Anaesthesia System **Envitec** 1000470 Flow Sensor **Envitec** SpiroQuant A+ Flow Sensor 0032 00 02 Flow Sensor Integral Process Integral Process 65025 Flow Sensor Flow Sensor **Integral Process** 8403735 Intersurgical 2777000 Flow Sensor Flow Sensor Intersurgical e-FlowSensor M-45-00-0005 Medacx Flow Sensor SpiroQuant A Flow Sensor Medacx 1000470 **Proact Medical** Flow Sensor

SpiroQuant A+

Flow Sensor

Reviewed by: Steve Nixon

Director - Viamed Ltd.

Date: 21st July 2017

**Proact Medical**