

Post Market Surveillance Review

4310003 Flowsensor SpiroTrue H. Box of 6

First sold: **8th May 2009**

Total sales to date: **2,700 boxes, across 19 countries**

Total failures: **2 boxes**

Total returns: **2 packs – failed calibration on Hamilton G5 ventilator, reported then returned to Viamed for credit 10/12/2009. Sensors supplied to Bluepoint Medical for evaluation 29/12/2009**

Customer: **Royal Sussex County Hospital SRS61983**

Total customer failures: **0.07%**

Applications:

Manufacturer	Description	
Draeger	Oxilog 2000 Plus (if disposable circuits used)	Ventilator
Draeger	Oxilog 3000 (if disposable circuits used)	Ventilator
Draeger	Oxilog 3000 Plus (if disposable circuits used)	Ventilator
Draeger	VentStar	Ventilator
Hamilton Medical	155362	Flow Sensor
Hamilton Medical	279331	Flow Sensor
Hamilton Medical	53120	Flow Sensor
Hamilton Medical	Amadeus	Ventilator
Hamilton Medical	C1	Ventilator
Hamilton Medical	C2	Ventilator
Hamilton Medical	G5	Ventilator
Hamilton Medical	Galileo	Ventilator
Hamilton Medical	H72	Ventilator
Hamilton Medical	H7200	Ventilator
Hamilton Medical	H7400 AccuTach	Flow Sensor
Hamilton Medical	MR1	Ventilator
Hamilton Medical	Raphael	Ventilator
Hamilton Medical	S1	Ventilator
Hamilton Medical	T1	Ventilator
Hamilton Medical	Veolar	Ventilator
Hamilton Medical	VR-1	Ventilator

Reviewed by: Steve Nixon

Director – Viamed Ltd.

Date: 21st July 2017