

Service Repair Sheet 69399

Contact Name Karl Green
Company/ Hospital Name Alexandra Hospital
Department Technical Services
Direct Phone 01562823424
General Phone 01527503030
Opera Account 00004295
Email karl.green1@nhs.net
Order Number 305684655
Date Received 07/Apr/2026
Booked in By Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN38687	1410000	PR02913A10	Foetal Simulator	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
		Qty: 1	1480000	
		Qty: 1	1430309	

V1000 has had transducer interface cushion replaced and calibration has been checked.
