

Service Repair Sheet SRS69334

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Sean Bunn

Luton & Dunstable Hospital

Clinical Engineering

01582718369

01582497042

00003490

Sean.Bunn@ldh.nhs.uk

15/Jan/2026

Robert Connor

Viamed

Quote

VIAMEDclean

Goods In Only

Decontamination
certificate provided
by customer ☒

Cleaned by Viamed,
if no declaration certificate
from customer ☒

Signed: _____

Date: _____

Goods Out Only

Cleaned by Viamed before
returning to customer

Signed: _____

Date: _____

Notes 14/Jan/2026 Catrin Hird

14/Jan/2026 Catrin Hird

Customer returning VM-2500-S, S/N: A1711160013

The unit needs recalibration, but they usually do this themselves but when checking the percentage of the CO2 was off, even after zeroing. Please review unit and confirm with customer before proceeding with any calibration or work.

15/Jan/2026 Robert Connor

Received 1 x VM-2500-S s/n A1711160013, with silicone cover and 4 x AA battery.

Ready For quote

CGreen 20.2.26

Repair Complete Signed

CGreen 14.4.26

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN38504	Capnograph	4410520	A1711160013	N

4440000 x 45mins @ £95 per hr. - Includes our time + Blueprints.
S/N, SRS, SRN

4480520 x 1 @ £160
SRS, SRN

4430201 x 1 @ £15.40
SRS, SRN

4430050 x 1 @ £2.85.
SRS, SRN

UPS x 1 @ £12.