

Service Repair Sheet SRS69408

Contact Name Ivana Krasna
 Company/ Hospital Name Chirana Medical a.s.
 Department Purchasing
 Position
 Direct Phone 421918714050
 General Phone 00421918714072
 Opera Account 00007190
 Email ikrasna@chirana.eu
 Order Number
 Date Received 21/Apr/2026
 Booked in By Robert Connor
 Main Company Viamed
 Type Return Warranty

VIAMEDclean	
Goods In Only	
Decontamination certificate provided by customer	<input checked="" type="checkbox"/>
Cleaned by Viamed, if no declaration certificate from customer	<input type="checkbox"/>
Signed: _____	
Date: _____	
Goods Out Only	
Cleaned by Viamed before returning to customer	
Signed: _____	
Date: _____	

Notes 14/Apr/2026 Kate Griffiths
 14/Apr/2026 Kate Griffiths
 Customer has 2 x 0110017 under warranty, s/n 259680 and 259681. They have checked them with 2 devices and it shows 0mv (close to)
 15/Apr/2026 Kate Griffiths
 UPDATE: 1ZR0E4566804174510 UPS tracking no. from customer
 21/Apr/2026 Robert Connor
 Received 2 x R17MED s/n 259680, 259681.

Ready For quote _____

Repair Complete Signed _____

 CGREEN 22.4.26

USD
 DIST.

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN38746	Oxygen Sensor	0110017	259680	Yes
SRN38747	Oxygen Sensor	0110017	259681	Yes

0140013 x 2 FOC
 Old S/N's SRS, SRN's

0110017 x 2 FOC
 New S/N's SRS, SRN's

PPUPS6 . 23x15x12cm.
 Add declared value to UPS