

Service Repair Sheet 69407

Contact Name Gail Lohfink
Company/ Hospital Name Dorset County Hospital
Department Clinical Engineering
Position Clinical Engineering Support Officer
Direct Phone 01305254118
General Phone 01305251150
Opera Account 00001430
Email gail.lohfink@dchft.nhs.uk
Order Number P249482
Date Received 15/Apr/2026
Booked in By Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN38720	1410000	PR065A18	Foetal Simulator	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
		Qty: 1	1480000	
		Qty: 1	1430309	

V1000 has had transducer interface cushion replaced and calibration has been checked.
