

Service Repair Sheet 69415

Contact Name David Watson
Company/ Hospital Name Darlington Memorial Hospital
Department Clinical Engineering Department
Position Clinical Engineering Technician
Direct Phone 01325743048
General Phone 01325380100
Opera Account 00001350
Email david.watson28@nhs.net
Order Number CDD0006592332
Date Received 27/Apr/2026
Booked in By Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN38751	0310030	VMD105	Tom Thumb	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
	Qty: 1	0380000		

Tom Thumb TT480 VMD105 has been stripped and cleaned then reassembled with new O rings. The two valves have been checked and reset before QA.

Repair	Ref	S/N	Equipment Type	
SRN38752	0310030	VMDH10	Tom Thumb	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
	Qty: 1	0380000		

Tom Thumb TT490 VMDH10 has been stripped and cleaned then reassembled with new O rings. The two valves have been checked and reset before QA.
