

Service Repair Sheet SRS69417

Contact Name Tracy Wang  
 Company/ Hospital Name Justec Shenzhen Co. Ltd  
 Department  
 Position  
 Direct Phone 008675526068930  
 General Phone 008675526068931  
 Opera Account 00006027  
 Email tracy.wang@justec.cn  
 Order Number  
 Date Received 29/Apr/2026  
 Booked in By Robert Connor  
 Main Company Viamed  
 Type Return Test Only

<b>VIAMEDclean</b>	
<b>Goods In Only</b>	<input checked="" type="checkbox"/>
Decontamination certificate provided by customer	
<b>Cleaned by Viamed, if no declaration certificate from customer</b>	<input type="checkbox"/>
Signed: _____	
Date: _____	
<b>Goods Out Only</b>	
Cleaned by Viamed before returning to customer	
Signed: _____	
Date: _____	

Notes From RS, just past warranty, but ok. Customer will return 2 x 0110081, S/N 241698 and 241696. Bought in March 2025, not under warranty. Pics/more info in email attached. 241698 - initial reading only 16. 241696 - approx 10 percent below normal

*Ryan ok'd to cover as warranty 05-26*

24/Apr/2026 Kate Griffiths  
 From RS, just past warranty, but ok. Customer will return 2 x 0110081, S/N 241698 and 241696. Bought in March 2025, not under warranty. Pics/more info in email attached.

241698 - initial reading only 16. 241696 - approx 10 percent below normal

27/Apr/2026 Kate Griffiths  
 UPDATE, on way - UPS 1ZAF71610497340354.

29/Apr/2026 Robert Connor  
 Received 2 x UFO 130-2 s/n 241696, 241698.

*USD Dist*

Ready For quote \_\_\_\_\_

Repair Complete Signed \_\_\_\_\_

*CGNOR 6.5.26*

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN38758	Oxygen Sensor	0110081	241696	2
SRN38759	Oxygen Sensor	0110081	241698	2

*6140013 x 2 fOC  
 old S/N, SRS, SRN*

*0110081 x 2 fOC  
 New S/N, SRS, SRN  
 customs value. \$166.50.*

*UPS x 1 fOC*

*DDP.*