

Service Repair Sheet 69411
Contact Name Adam Wharton
Company/ Hospital Name North Cumbria University Hospitals NHS Trust
Department EBME
Direct Phone 01228814671
General Phone 01228523444
Opera Account 00000980
Email adam.wharton@ncic.nhs.uk
Order Number RNNN400318308
Date Received 23/Apr/2026
Booked in By Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN38749	1410000	PR01225A15	Foetal Simulator	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
		Qty: 1	1430300	
		Qty: 0.5	1440000	

The V1000 has had a PCB failure.

I have replaced the PCB and issued a new calibration certificate.
