

Service Repair Sheet SRS69411

Contact Name Adam Wharton
 Company/ Hospital Name North Cumbria University Hospital
 Department EBME
 Position
 Direct Phone 01228814671
 General Phone 01228523444
 Opera Account 00000980
 Email adam.wharton@ncic.nhs.uk
 Order Number
 Date Received 23/Apr/2026
 Booked in By Robert Connor
 Main Company Viamed
 Type Return Quote

VIAMEDclean	
Goods In Only	<input checked="" type="checkbox"/>
Decontamination certificate provided by customer	
Cleaned by Viamed, if no declaration certificate from customer	<input checked="" type="checkbox"/>
Signed: <u>RC</u>	
Date: <u>23/4/26</u>	
Goods Out Only	
Cleaned by Viamed before returning to customer	
Signed: _____	
Date: _____	

Notes 20/Apr/2026 Aqib Majeed

20/Apr/2026 Aqib Majeed

They have a v1000 device SN: PR01225A15 which appears to be no longer emitting a signal, they have replaced batteries but still not resolved.

23/Apr/2026 Robert Connor

Received 1 x V1000 s/n PR01225A15, with 4 x AA battery and blue fabric carry case.

Out of warranty, changed to quote.

Please remind customer that a declaration of decontamination status is required, and that sending paperwork with a repair will help identify incoming parcels and ensure their repairs are dealt with in a timely manner.

Ready For quote

CGNEEN 27.4.26

Repair Complete Signed

CGNEEN 7.5.26

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN38749	Foetal Simulator	1410000	PR01225A15	N

1440000 x 1 @ 0.5hr. @ £95 per hr = £47.50.
 S/N SRS, SRN

1430300 x 1 @ £97.50

~~S/N~~, SRS, SRN

UPS x 1 @ £12.
~~14400~~