

Service Repair Sheet 69427

Contact Name	Purchasing
Company/ Hospital Name	Eden Medical (UK) Ltd
Direct Phone	01314406500
General Phone	01314406500
Opera Account	00012178
Email	info@edenmedical.co.uk
Date Received	11/May/2026
Booked in By	Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN38770	1410000	PR02643A11	Foetal Simulator	Complete - Repaired Time :0 Hour(s)

Parts Replaced

Qty: 1 1480000

Qty: 1 1430309

V1000 has had transducer interface cushion replaced and calibration has been checked.
