



MOD Obsolescence review Microstim_DB3

Catrin Hollings <catrin.hollings@viamed.co.uk>
To: "Leidos-TVS-David.Ford" <david.ford121@teamleidos.mod.uk>
Cc: Kate Griffiths <kate.griffiths@viamed.co.uk>
Bcc: Main Account <viamedinbox@gmail.com>

12 May 2026 at 11:36

Hi David

Thank you for your email, please find attached the completed table below. Answers for previously asked questions remain the same:

Although the 7 years from the date of discontinuation of the DB3 will lapse shortly, we are happy to extend for another year and this will be reviewed on an annual basis.

Information/Name/MPN	p/n 2510000 - Microstim DB3
Confirm that " Viamed" is the Original Manufacture	Viamed is the Original Manufacturer
Provide the OEM MHRA Reference Number (As shown on the MHRA PARD)	n/a
Provide the PARD GMDN for the device	n/a
Is the equipment available to purchase? (If possible please provide current price ex VAT)	No, it was discontinued 17th May 2019
Do you have an end of support date or obsolescent date for the equipment?	17th May 2027
Do you have a planned end of production for this model?	Production discontinued 17th May 2019
If no longer available please provide an end of life letter.	Attached
Please provide an Up to date list of Ancillaries (if Applicable)	See attached spare parts price list

Provide the latest list of spare parts and prices for the equipment (If applicable)	See attached spare parts price list
Provide the latest version of the user and service manual for the equipment	Attached
Provide the latest material safety data sheet (Battery Safety Data Sheet, UN38.3)	Attached
Please provide a DofC for this equipment.	Attached
Provide a UKCA and/or CE certification for the equipment	Manufacturer no longer applies CE mark to this product and will not be applying UKCA
Please Provide the latest EC certificate	See Dof C
Please provide the latest QMS certificate (ISO 9001& ISO 13485)	Attached
What software version should the equipment have installed? (if applicable)	n/a
Have there been any FSN modifications we need to be aware of?	No
Can clinical training be provided for this equipment? Is this certified? How long is the certificate valid for? Can a cost please be provided?	Training not available
Is a clinical refresher training course available?	Training not available
Can technical training be provided to allow in house engineers to maintain this equipment? Is this certified? How long is the certificate valid for?	Training not available
If a technical course is available what is the course length? What is the maximum number of attendees? Can a cost please be provided?	Training not available
Is a technical refresher training course available?	Training not available

Please be aware that the MySign O has also now been discontinued.

Should you need any further information please let me know.

Kind regards

Catrin Hollings
Marketing
Viamed Ltd.

<http://www.viamed.co.uk>

Email: catrin.hollings@viamed.co.uk

Tel: +44 (0)1535 634 542

Fax: +44 (0)1535 635 582

Working Hours: Tuesday - Thursday 8.30am - 2.30pm

Check out our product portfolio at:

www.viamed.co.uk/productportfolio

Twitter: twitter.com/ViamedLtd

Facebook Page: Search for Viamed Ltd

Company registered in England, No. 01291765.

E-mail Disclaimer

GDPR / Privacy Policy available at www.viamed.co.uk/policy

The information in this email is confidential and may be legally privileged.

It is intended solely for the addressee. Access to this email by anyone else is unauthorized.

If you are not the intended recipient, any disclosure, copying, distribution or any action taken or omitted to be taken in reliance on it, is prohibited and may be unlawful.


If you feel you have received this email in error or would like to be removed from our mailing list, please reply with delete in the subject line or call +44 (0)1535 634542

[Quoted text hidden]

4 attachments

 **GBP Price List - Microstim DB3 Spare Parts ID218521.pdf**
88K

 **Microstim MkIII EC Declaration of Conformance 2510000 ID28847.pdf**
277K

 **Viamed ISO13485-2016 Certificate ID176874.pdf**
449K

 **Notice of Discontinuation - Microstim DB3.pdf**
235K