

Service Repair Sheet 69416
Contact Name Clay Leary
Company/ Hospital Name Sage Services Group
Position Service Operations Director
Direct Phone 8645991922
General Phone 8432845844
Opera Account 00006727
Email cleary@sageservicesgroup.com
Date Received 29/Apr/2026
Booked in By Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN38760	1410000	PR02516A10	Foetal Simulator	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
	Qty: 1	1480000		
	Qty: 1	1430309		

V1000 has had transducer interface cushion replaced and calibration has been checked.
