

Service Repair Sheet 69406

Contact Name Aneta Szubarga
Company/ Hospital Name Euromed Medical Solution Sp. z o.o
Direct Phone 0048618439059
General Phone 504 078 773
Opera Account 00006582
Email a.szubarga@euromed.net.pl
Date Received 13/May/2026
Booked in By Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN38782	1410000	PR02724A13	Foetal Simulator	Complete - Repaired Time :0 Hour(s)

Parts Replaced

Qty: 1 1480000

Qty: 1 1430309

V1000 has had transducer interface cushion replaced and calibration has been checked.
