

## **Product Changes**

| Date        | Change to product  |
|-------------|--|
| Dec 1994    | Outlets changed to 15mm  |
| July 1995   | Added the subtracted restrictor in O2 tubing inlet ( used to prevent EBME;destroying during acceptance tests)  |
| June 1996   | Suppliers changed  |
| Oct 1996    | Altered safety valve into a precision valve by removing the ball   |
| Nov 1996    | Grease changed ( supplier reasons)   |
| December 96 | Discussions with Drager on modification extended block by 32mm Increase the bore to accomodate two flow meters |
|             |  |
|             |  |
|             |  |
|             |  |
|             |  |
|             |  |
|             |  |
|             |  |
|             |  |
|             |  |
|             |  |
|             |  |
|             |  |