

Certificate of Calibration

Certificate number: 1780475843

V1000 Foetal Heart Simulator

Part Number: 1410000

Serial number: PR03920A14

This certificate reports recorded values for the instrument specified above.

The output rates were measured in accordance with procedure:
VM3COP55.01.

Equipment used: Sonicaid FM6 foetal monitor (CE051)

TTi TF930 3GHz counter (CE185/CE216) -
calibration traceable to UKAS calibration standards.

| V1000 BPM setting | Period values stated in milliseconds | | | | Result PASS/FAIL |
|----------------------|--------------------------------------|----------------|----------------|---------------|---------------------|
| | Target | Lower limit | Upper limit | Actual | |
| 60 | 1000.00 | 990.00 | 1010.00 | 995.25 | PASS |
| 90 | 666.70 | 660.03 | 673.37 | 667.08 | PASS |
| 120 | 500.00 | 495.00 | 505.00 | 501.11 | PASS |
| 150 | 400.00 | 396.00 | 404.00 | 400.99 | PASS |
| 180 | 333.30 | 329.97 | 336.63 | 333.34 | PASS |
| 210 | 285.70 | 282.84 | 288.25 | 286.44 | PASS |

Name: Phil Crossley

Date: 3rd June 2026

Signed: _____