

Service Repair Sheet SRS69419

Contact Name	Sudath Weeraman
Company/ Hospital Name	Prime Lanka Associate
Department	
Position	
Direct Phone	9410718732285
General Phone	
Opera Account	00006714
Email	sudathweeraman17@gmail.com
Order Number	
Date Received	26/May/2026
Booked in By	Robert Connor
Main Company	Viamed
Type Return	Warranty

VIAMEDclean	
Goods In Only	<input checked="" type="checkbox"/>
Decontamination certificate provided by customer	
Goods Out Only	<input type="checkbox"/>
Cleaned by Viamed, if no declaration certificate from customer	
Signed: _____	
Date: _____	
Goods Out Only	
Cleaned by Viamed before returning to customer	
Signed: _____	
Date: _____	

Notes 27/Apr/2026 Kate Griffiths
 27/Apr/2026 Kate Griffiths
 will send in 1 x MAX-250E 0110429, s/n: LH45899010 - issue is high oxygen value (he says Valve but I assumed he meant value..) Purchased Feb 2026
 26/May/2026 Robert Connor
 Received 1 x MAX-250E s/n LH45899010

Ready For quote _____

Repair Complete Signed _____

CGreen 26/5/26

*Exw.
Ship with next order*

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN38799	Oxygen Sensor	0110429	LH45899010	Yes

*0140013 x 1 FOC
old S/N, SRS, SRN*

*0110429 x 1 FOC
new S/N, SRS, SRN*

*Exw - ship with next order.
sky Cargo*