

Certificate of Calibration

Certificate number: 1780572853

Tom Thumb Infant Resuscitator.

Part Number: 0310030

Serial number: VMEF03

This certificate reports recorded values for the instrument specified above.

The output rates were measured in accordance with Tom Thumb service manual.

Equipment used: Digitron Digital Manometer P200 (CE178)
Digitron Digital Manometer P200 (CE091)
Flow Meter (CE130)
Snoop Liquid

| Test Equipment | Test | Specification | Reading | Pass/Fail |
|---------------------------|---|---|---------|-----------|
| Snoop liquid | Check all ports / connections for leaks | No bubbling | N/A | PASS |
| Digital Manometer (CE178) | Adjustable Valve: @ 10 lpm | Minimum: <7.8 mbar | 2.8 | PASS |
| Digital Manometer (CE178) | Adjustable Valve: @ 10 lpm | Maximum: 42.1 ±1 mbar | 41.9 | PASS |
| Digital Manometer (CE178) | Precision Valve: @ 10 lpm | Maximum: 44.1 ±1 mbar | 43.3 | PASS |
| Digital Manometer (CE178) | Pressure Gauge Test @30cmH ₂ O | Pressure: 30cmH ₂ O ±1 (28.4 – 30.4mbar) | 29.2 | PASS |

Name: Phil Crossley

Date: 4th June 2026

Signed: _____