

Service Repair Sheet 69443

Contact Name Adam Allan
Company/ Hospital Name Royal Alexandra Hospital
Department Medical Physics
Position Associate Practitioner Technologist
Direct Phone 01413146637
General Phone 0141 887 9111
Opera Account 00004080
Email adam.allan@nhs.scot
Order Number GMPA16641455
Date Received 08/Jun/2026
Booked in By Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN38810	1410000	PR03102A10	Foetal Simulator	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
		Qty: 1	1480000	
		Qty: 1	1430309	

V1000 has had transducer interface cushion replaced and calibration has been checked.
