

Service Repair Sheet 69438

Contact Name Clay Harper
Company/ Hospital Name Sage Services Group
Direct Phone 8432845844
General Phone 8432845844
Opera Account 00006261
Email harper@sageservicesgroup.com
Date Received 28/May/2026
Booked in By Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN38803	1410000	PR01883A18	Foetal Simulator	Complete - Repaired Time :0 Hour(s)

Parts Replaced

Qty: 1 1480000

Qty: 1 1430309

V1000 has had transducer interface cushion replaced and calibration has been checked.
