



Cathy Green &lt;viamed.cathy.green@gmail.com&gt;

---

**SRS6944**

1 message

---

**Cathy Green** <cathy.green@viamed.co.uk>

10 June 2026 at 15:34

To: carl@generaldiagnostic.co.uk

Bcc: Main Account &lt;viamedinbox@gmail.com&gt;

Dear Carl

We have received your automotive sensors R-17A-LVG, two stated as out of box failures and one no out put.  
We have fully tested the units, please see results below:

S/N:V012144 tested, no fault found, tested to within specification. Please advise if you would like me to return as no fault found.

S/N:V102251 No output, will replace under warranty.

S/N:V102439 No output, will replace under warranty.

Please do not hesitate to contact me further.

--

Kind regards

Cathy Green  
Goods Out / Stock Control Supervisor  
Viamed Ltd.

<http://www.viamed.co.uk>

Email: [cathy.green@viamed.co.uk](mailto:cathy.green@viamed.co.uk)

Tel: 44 (0)1535 634 542

Fax: 44 (0)1535 635 582

Twitter: [twitter.com/ViamedLtd](https://twitter.com/ViamedLtd)

Facebook Page: Search for Viamed Ltd

Company registered in England, No. 01291765.

E-mail Disclaimer

-----

GDPR / Privacy Policy available at [www.viamed.co.uk/policy](http://www.viamed.co.uk/policy)

The information in this email is confidential and may be legally privileged.

It is intended solely for the addressee. Access to this email by anyone else is unauthorized.

If you are not the intended recipient, any disclosure, copying, distribution or any action taken or omitted to be taken in reliance on it, is prohibited and may be unlawful.

If you feel you have received this email in error or would like to be removed from our mailing list, please reply with delete in the subject line or call 44 (0)1535 634542

-----