

Service Repair Sheet SRS69422

Contact Name Engel Bejarin
 Company/ Hospital Name Kings College Hospital (use this o
 Department
 Position
 Direct Phone 02032999000
 General Phone 02032997878
 Opera Account 00002990
 Email e.bejarin@nhs.net
 Order Number
 Date Received 05/May/2026
 Booked in By Robert Connor
 Main Company Viamed
 Type Return Warranty

VIAMEDclean	
Goods In Only	<input checked="" type="checkbox"/>
Decontamination certificate provided by customer	
Cleaned by Viamed, if no declaration certificate from customer <input checked="" type="checkbox"/>	
Signed: _____	
Date: _____	
Goods Out Only	
Cleaned by Viamed before returning to customer	
Signed: _____	
Date: _____	

Notes 28/Apr/2026 Kate Griffiths

28/Apr/2026 Kate Griffiths

Engel rang in, the switch on his MAXO2 ME does not respond. The s/n is KH63699017. Warranty valid until Sept. 2026. Will ask for cert of decontamination/batteries removed SRS on all paperwork/cable and sensor included

05/May/2026 Robert Connor

Received 1 x MAXO2ME s/n KH63699017, with coiled cable ID2545193, MAX550E s/n KE96199102, with flow diverter, battery door missing kickstand.

Ready For quote

Repair Complete Signed

CGreen 11-6-26

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN38761	Oxygen Monitor	0111263	KH63699017	Yes
SRN38762	Oxygen Monitor Cable	0131700	Tracking4	Yes
SRN38763	Oxygen Sensor	0110452	KE96199102	Yes

0140020 x 1 FOC
S/N, SRS, SRN
Repaired under warranty

0140025 x 1 FOC
no fault
S/N, SRS, SRN

0140011 x 1 FOC
S/N, SRS, SRN
no fault

UPS x 1 FOC.