

Service Repair Sheet 69433
Contact Name Barakat Suleiman
Company/ Hospital Name Wythenshawe Hospital
Position Medical Engineering
Direct Phone 01612912019
General Phone 0161 291 2019
Opera Account 00003640
Email barakat.suleiman@mft.nhs.uk
Order Number 000101_100029392
Date Received 26/May/2026
Booked in By Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN38801	1410000	PR03746A12	Foetal Simulator	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
		Qty: 1	1480000	
		Qty: 1	1430309	

V1000 has had transducer interface cushion replaced and calibration has been checked.
