

Service Repair Sheet SRS69444

Contact Name Carl Wodke  
 Company/ Hospital Name General Diagnostic Limited  
 Department  
 Position  
 Direct Phone 01455617888  
 General Phone 01455617888  
 Opera Account 00020053  
 Email carl@generaldiagnostic.co.uk  
 Order Number  
 Date Received 08/Jun/2026  
 Booked in By Robert Connor  
 Main Company Viamed  
 Type Return Warranty

<b>VIAMEDclean</b>	
<b>Goods In Only</b>	
Decontamination certificate provided by customer	<input type="checkbox"/>
Cleaned by Viamed, if no declaration certificate from customer	<input type="checkbox"/>
Signed: _____	
Date: _____	
<b>Goods Out Only</b>	
Cleaned by Viamed before returning to customer	
Signed: _____	
Date: _____	

Notes 03/Jun/2026 Kate Griffiths  
 03/Jun/2026 Kate Griffiths  
 has 3 x 0110157 under warranty. S/N: V102251, V102144 and V102439. 2 failed out of box. V102439 0 voltage output  
 08/Jun/2026 Robert Connor  
 Received 3 x R17ALVG s/n V102144, V102251, V102439.

Ready For quote

Repair Complete Signed

CGNEEN 12-6-26

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN38807	Oxygen Sensor	0110157	V102144	✓
SRN38808	Oxygen Sensor	0110157	V102251	Yes
SRN38809	Oxygen Sensor	0110157	V102439	Yes

should be warranty of 8/6/26

0140011 x 1 FOC  
 S/N, SRS, SRN

0140013 x 2 FOC  
 old S/N's, SRS, SRN's

0110157 x 2 FOC  
 new S/N's SRS, SRN's

GPS x 1 FOC