

Service Repair Sheet 69446  
Contact Name Claire Willis  
Company/ Hospital Name Medway Maritime Hospital  
Department Clinical Engineering  
Position Clinical Engineering Inventory and Systems Assistant  
Direct Phone 01634976451  
General Phone 01634830000  
Opera Account 00001770  
Email c.willis2@nhs.net  
Order Number 270002207  
Date Received 09/Jun/2026  
Booked in By Robert Connor

---

Repair	Ref	S/N	Equipment Type	
SRN38875	1410000	PR02094A30	Foetal Simulator	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
		Qty: 1	1480000	
		Qty: 1	1430309	

V1000 has had transducer interface cushion replaced and calibration has been checked.

---