

Service Repair Sheet SRS69439
Contact Name
Company/ Hospital Name
Department
Position
Direct Phone
General Phone
Opera Account
Email
Order Number
Date Received
Booked in By
Main Company
Type Return

Rob Pochwyt
Bar O2 Limited

01785338133
CID11243
support@baro2.com

03/Jun/2026
Robert Connor
Viamed
Warranty

VIAMEDclean

Goods In Only

Decontamination
certificate provided
by customer ☐

Cleaned by Viamed,
if no declaration certificate
from customer ☐

Signed: _____
Date: _____

Goods Out Only

Cleaned by Viamed before
returning to customer

Signed: _____
Date: _____

Notes 28/May/2026 Sophie Lines

28/May/2026 Sophie Lines

Has 2x faulty Ultramax Oxygen Analysers. Serial numbers KA21311020 and KA21311017. Should still be covered under warranty. I have asked for further details of faults and instructions for return.

28/May/2026 Sophie Lines

Reply from Rob regarding faults: one of the analysers will not turn on and the other analyser will not calibrate at all and also on one of the analysers, it feels like the one of the buttons is not reading or reacting when pressed.

03/Jun/2026 Robert Connor

Received 2 x UltramaxO2 s/n KA21311017, KA21311020, with 2 x tubing and 2 x blue silicone cover.

Ready For quote

Repair Complete Signed

CC Connor 16.6.26

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN38804	Oxygen Analyser	0111265	KA21311017	Yes
SRN38805	Oxygen Analyser	0111265	KA21311020	Yes

0140050 x 2 FOC

SIN's, SRS, SRN's
faulty, see line below for replacements

0111265 x 2 FOC

New SIN's, SRS, SRN's
warranty Replacement

UPS x 1 FOC