

Service Repair Sheet 69455
Contact Name Simon Pugh
Company/ Hospital Name Princess Royal Hospital
Department Medical Engineering Services
Position Medical Engineer
Direct Phone 01952 641222 EXT 4262
General Phone 01743261149
Opera Account 00005117
Email simon.pugh3@nhs.net
Date Received 17/Jun/2026
Booked in By Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN38882	1410000	PR0773A14	Foetal Simulator	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
	Qty: 1	1480000		
	Qty: 1	1430309		

V1000 has had transducer interface cushion replaced and calibration has been checked.
