

Service Repair Sheet SRS69453

Contact Name Paul Denyer
 Company/ Hospital Name Crawley Hospital
 Department EME Department
 Position EME Services Technician
 Direct Phone 01293600300
 General Phone 01737768511
 Opera Account 00001300
 Email paul.denyer@nhs.net
 Order Number
 Date Received 18/Jun/2026
 Booked in By Robert Connor
 Main Company Viamed
 Type Return Quote

VIAMEDclean

Goods In Only

Decontamination
 certificate provided
 by customer ☒

Cleaned by Viamed,
 if no declaration certificate
 from customer ☐

Signed: _____

Date: _____

Goods Out Only

Cleaned by Viamed before
 returning to customer

Signed: _____

Date: _____

Notes They would like to send in a DB3, s/n H0002232 as the intensity control knob/signal strength variation control knob is broke. I asked if they would prefer just to purchase the knob. Paul replied that the knob is well and truly fixed in place and he can hear what it controls turning inside the unit. He prefers to send it in.

15/Jun/2026 Kate Griffiths

They would like to send in a DB3, s/n H0002232 as the intensity control knob/signal strength variation control knob is broke. I asked if they would prefer just to purchase the knob. Paul replied that the knob is well and truly fixed in place and he can hear what it controls turning inside the unit. He prefers to send it in.

18/Jun/2026 Robert Connor

Received 1 x DB3 s/n H0002232, with patient lead.

Ready For quote

CCGreen 22.6.26

Repair Complete Signed

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN38884	Microstim DB3	2510000	H0002232	

2540010 x 1 @ £70

S/N, SRS, SRN

UPS x 1 @ £12