

Service Repair Sheet 69447

Contact Name	Stephane Palmans
Company/ Hospital Name	Arseus Medical Group NV
Department	Customer Service
General Phone	3280076998
Opera Account	00007530
Email	contractadministration@arseus-medical.be
Order Number	26-23599
Date Received	15/Jun/2026
Booked in By	Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN38881	1410000	PR02586A13	Foetal Simulator	Complete - Repaired Time :0 Hour(s)

Parts Replaced

Qty: 1	1480000
Qty: 1	1430309

V1000 has had transducer interface cushion replaced and calibration has been checked.
