

Service Repair Sheet 69449

Contact Name	Ferdinand De Guzman
Company/ Hospital Name	University College Hospital
Department	Medical Physics and Biomedical Engineering
Position	Biomedical Engineer
Direct Phone	02034476255
General Phone	0845 1555000/02034 567890
Opera Account	00003420
Email	f.deguzman@nhs.net
Date Received	15/Jun/2026
Booked in By	Robert Connor

Repair	Ref	S/N	Equipment Type	Under Warranty
SRN38879	1410000	PR03812A13	Foetal Simulator	Complete - Repaired
Parts Replaced				Time :0 Hour(s)
	Qty: 1	1480000		
	Qty: 1	1430309		

V1000 has had transducer interface cushion replaced and calibration has been checked.
