

Service Repair Sheet SRS69436

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Lee Harper

Pinderfields General Hospital

Medical Physics Department

01924541050

01924541050

00005190

lee.harper@nhs.net

26/May/2026

Robert Connor

Viamed

Quote

VIAMEDclean

Goods In Only

Decontamination
certificate provided
by customer ☒

Cleaned by Viamed,
if no declaration certificate
from customer ☒

Signed: _____

Date: _____

Goods Out Only

Cleaned by Viamed before
returning to customer

Signed: _____

Date: _____

Notes 20/May/2026 Kate Griffiths

20/May/2026 Kate Griffiths

Their TOF3D, S/N: T3D202101022 "powers on and gives the splash screen. It then has nothing else on screen but a circle in the top right corner. Faulty main cable?"

Will ask for cert of decont/no batteries/all accessories/quote SRS on all paperwork

26/May/2026 Robert Connor

Received 1 x TOF3D s/n T3D202101022, with cable and patient lead, blemish on upper left corner of screen.

Ready For quote

Repair Complete Signed

CGreen 29-6-26

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN38802	Microstim	2510091	T3D202101022	

2580100 x 1 FOC

SIN, SRS, SEN

No fault found

GPS x 1 FOC.