

### **Sterilisation.**

The Tom Thumb Unit does not normally require sterilisation (except for the mouth-piece and tubing of disposables – these are not supplied).

Apart from the “O-rings” the unit is chromed brass, so it can be dismantled and “autoclaved” if required.