



Cathy Green <viamed.cathy.green@gmail.com>

SRS69463

1 message

Cathy Green <cathy.green@viamed.co.uk>
To: Stephen Machin <stephen@prosol.co.uk>
Bcc: Main Account <viamedinbox@gmail.com>

8 July 2026 at 09:03

Dear Stephen

Good Moring

We have tested the 10 sensors that you have sent in for testing.
We have found that S/N P107864 (0110365) and S/N: P107729 (0110365) are both faulty, which we will replace under warranty
S/N:P107873 (0110365) has no output to be replaced under warranty,

The other 7 sensors are all testing with in specification, how would you like to proceed with these, would you like them sending back as no fault found.

I am out of the office for the rest of the day, but will be back in work on Thursday 09/07/26.

--

Kind regards

Cathy Green
Goods Out / Stock Control Supervisor
Viamed Ltd.

<http://www.viamed.co.uk>
Email: cathy.green@viamed.co.uk
Tel: 44 (0)1535 634 542
Fax: 44 (0)1535 635 582

Twitter: twitter.com/ViamedLtd
Facebook Page: Search for Viamed Ltd

Company registered in England, No. 01291765.

E-mail Disclaimer

GDPR / Privacy Policy available at www.viamed.co.uk/policy

The information in this email is confidential and may be legally privileged.

It is intended solely for the addressee. Access to this email by anyone else is unauthorized.

If you are not the intended recipient, any disclosure, copying, distribution or any action taken or omitted to be taken in reliance on it, is prohibited and may be unlawful.

If you feel you have received this email in error or would like to be removed from our mailing list, please reply with delete in the subject line or call 44 (0)1535 634542
