

RIGHTINGTON, WIGAN AND LEIGH TEACHING HOSPITALS
NHS FOUNDATION TRUST

MEDICAL EQUIPMENT REPAIR REQUISITION

Requisition No. **62736** Call Ref. No.
Equipment No. **042816**

Equipment Description **TOM THUMB TT480**

Hospital/Site **RAE** Ward/Dept **MATERNITY**

Location of equipment for repair (e.g. Sluice)

Fault **DIAL GESS STUCK ON 50 WHEN TESTED WHEN COES PAST 50 DOES NOT RETURN TO ZERO**

Potential Infection Hazard

Has the equipment been in contact with the following infection?

Please tick as appropriate.

Blood/Body Fluid

Yes ☐

No ☒

Faeces/Urine

Yes ☐

No ☒

Sputum

Yes ☐

No ☒

M.R.S.A

Yes ☐

No ☒

Initial Preparation

Has the surface been washed with detergent and water?

Yes ☒

No ☐

If any other infection control precautions are needed liase with Infection Control Department.

Date **07/07/76** Print Name **Chen Daniey**

Medical Electronics Department Use Only

Prior to Repair Equipment Decontaminated Yes ☐ No ☐

Equipment Repaired Yes ☐ No ☐

Post Repair Equipment Decontaminated Yes ☐ No ☐

Comments **SURFACE WIPE DOWN WITH**

SAOX CCORN 70%

Date **07/07/76** Print Name **Chen Daniey**

If this form is not completed or the equipment not cleaned it will **NOT** be repaired or serviced, it will be returned to the point of origin.