

Service Repair Sheet SRS69465

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Owen Daniels

Royal Albert Edward Infirmary

Medical Electronics

Medical Engineering Technician

01942822002

01942822002

00005363

Owen.Daniels@wwl.nhs.uk

09/Jul/2026

Robert Connor

Viamed

Quote

VIAMEDclean

Goods In Only

Decontamination
certificate provided
by customer ☒

Cleaned by Viamed,
if no declaration certificate
from customer ☐

Signed: _____

Date: _____

Goods Out Only

Cleaned by Viamed before
returning to customer

Signed: _____

Date: _____

Notes 06/Jul/2026 Kate Griffiths

06/Jul/2026 Kate Griffiths

They have a Tom Thumb TT480, s/n: 0401459. It is constantly sticking at 50 cm H2O. They will send it in. At the same time, please could their device be serviced? Last serviced 2022 as they do this in house.

09/Jul/2026 Robert Connor

Received 1 x tom thumb s/n 0401459, no rail clamp attached.

Ready For quote

CCNEEN 14.7.26

Repair Complete Signed

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN38915	Tom Thumb	0310030	0401459	N

0380000 x 1 @ £98

S/N, SRS, SRN

0330203 x 1 @ £133.30

SRS, SRN

0330212 x 1 @ £3.20

SRS, SRN

UPS x 1 @ £12.