

Calibration Certificate

The VN202 Oxygen analyser has been tested for accuracy by using a UKAS certified voltage source.

Calibrated voltage source	Microcal 1048
ISO calibration index number	CE226
Serial number of the Microcal	2435125
Certificate number of the Microcal	1978771
Serial number of the VN202	01343
Certificate number of the VN202	1784041864

Voltage input:	Target reading:	VN202 reading:
5.0mV	10%	9.9%
10.45mV	20.9%	20.8%
10.5mV	21%	20.9%
12.5mV	25%	24.9%
25.0mV	50%	49.9%
37.5 mV	75%	74.9%
50.0mV	100%	100%

Tested by: M.Green

Date: 14/07/2026

Signed: