

Contact Name	Paul Denyer
Company/ Hospital Name	Crawley Hospital
Department	EME Department
Position	EME Services Technician
Direct Phone	01293600300
General Phone	01737768511
Opera Account	00001300
Email	paul.denyer@nhs.net
Order Number	
Date Received	18/Jun/2026
Booked in By	Robert Connor
Main Company	Viamed
Type Return	Quote

VIAMEDclean
Goods In Only
Decontamination
certificate provided
by customer ☒

Cleaned by Viamed,
if no declaration certificate
from customer ☐
Signed: _____
Date: _____

Goods Out Only
Cleaned by Viamed before
returning to customer
Signed: _____
Date: _____

Notes They would like to send in a DB3, s/n H0002232 as the intensity control knob/signal strength variation control knob is broke. I asked if they would prefer just to purchase the knob. Paul replied that the knob is well and truly fixed in place and he can hear what it controls turning inside the unit. He prefers to send it in.

15/Jun/2026 Kate Griffiths

They would like to send in a DB3, s/n H0002232 as the intensity control knob/signal strength variation control knob is broke. I asked if they would prefer just to purchase the knob. Paul replied that the knob is well and truly fixed in place and he can hear what it controls turning inside the unit. He prefers to send it in.

18/Jun/2026 Robert Connor

Received 1 x DB3 s/n H0002232, with patient lead.

Ready For quote

CGreen 22.6.26

Repair Complete Signed

CGreen 17.7.26

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN38884	Microstim DB3	2510000	H0002232	

2540010 x 1 @ £70

S/N, SRS, SRN

UPS x 1 @ £12